



NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

Product: Anti-B

Lot No. 66013

Source of Material: Monoclonal antibody

Mfg. Date 15 JAN 23

Volume 10 mL, 50 mL

Exp. Date 15 JUL 24

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed yellow No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	B cell \geq 1:256 A ₁ B cell \geq 1:128	Complied
5. Specificity test	B cell = 4+ A ₁ B cell = 4+ A ₁ cell = Negative O cell = Negative	Complied
6. Avidity test	B cell < 6 sec A ₁ B cell < 6 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun Date 27/04/2023

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society