



NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

Product: Anti-B

Lot No. 66023

Source of Material: Monoclonal antibody

Mfg. Date 01 JUN 23

Volume 10 mL

Exp. Date 01 DEC 24

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed yellow No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	B cell \geq 1:256 A ₁ B cell \geq 1:128	Complied
5. Specificity test	B cell = 4+ A ₁ B cell = 4+ A ₁ cell = Negative O cell = Negative	Complied
6. Avidity test	B cell < 6 sec A ₁ B cell < 6 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun Date 24/08/2023

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society