



# NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

## CERTIFICATE OF ANALYSIS

**Product: Anti-B**

**Lot No. 66045**

**Source of Material: Monoclonal antibody**

**Mfg. Date 18 DEC 23**

**Volume 10 mL**

**Exp. Date 18 JUN 25**

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed yellow No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	B cell $\geq$ 1:256 A <sub>1</sub> B cell $\geq$ 1:128	Complied
5. Specificity test	B cell = 4+ A <sub>1</sub> B cell = 4+ A <sub>1</sub> cell = Negative O cell = Negative	Complied
6. Avidity test	B cell < 6 sec A <sub>1</sub> B cell < 6 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun Date 25/05/2023

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society