



ศูนย์บริการโลหิตแห่งชาติ
สภากาชาดไทย

NATIONAL BLOOD CENTRE
THAI RED CROSS SOCIETY

National Blood Centre
Thai Red Cross Society
Henri dunant Rd. Bangkok 10330
Tel (02) 2513111.
Fax (02) 2555558

CERTIFICATE OF ANALYSIS

REAGENT: Anti-Di^a (Human MoAb, IgM)
LOT NO: 65010
EXPIRE DATE: 8 AUG 23
SOURCE OF MATERIAL: Human monoclonal antibody
PRODUCT APPEARANCE: Clear Liquid
PRODUCT CHARACTERISTICS: Lot specific data (titer) are available
at NBC on request
MIN REQUIREMENT: Titer with Di(a+) cells \geq 1:4
SPECIFICITY: No other antibody demonstrable than the indicated
specificity
MICROBIOLOGICAL STATE: Terminal filtered to 0.2 μ m
and microbial count $<$ 1 CFU/mL
STORAGE: Refrigerate of +2 °C to +8 °C

SIGNED.....*S. Marking*.....DATE.....*23/08/23*.....

Quality officer