



# NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

## CERTIFICATE OF ANALYSIS

**Product:** Anti-M

**Lot No.** 66020

**Source of Material:** Monoclonal antibody

**Mfg. Date** 14 NOV 23

**Volume** 10 mL

**Exp. Date** 14 MAY 25

Test Items	Specifications	Results
1. Appearance	Clear liquid No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	MN cell $\geq$ 1:4	Complied
5. Specificity test	MN cell $\geq$ 3+ NN cell = Negative	Complied
6. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun Date 26/12/2023

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society