



NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

Product: Anti-Mi^a

Lot No. 67010

Source of Material: Monoclonal antibody

Mfg. Date 15 FEB 24

Volume 5 mL

Exp. Date 15 AUG 25

Test Items	Specifications	Results
1. Appearance	Clear liquid No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	Mi(a+) cell \geq 1:4	Complied
5. Specificity test	Mi(a+) cell \geq 2+ Mi(a-) cell = Negative	Complied
6. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun Date 12/03/2024

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society