



# NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

## CERTIFICATE OF ANALYSIS

Product: Anti-N

Lot No. 65020

Source of Material: Monoclonal antibody

Mfg. Date 25 AUG 22

Volume 10 mL

Exp. Date 25 FEB 24

Test Items	Specifications	Results
1. Appearance	Clear liquid No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	MN cell $\geq$ 1:4	Complied
5. Specificity test	MN cell $\geq$ 3+ MM cell = Negative	Complied
6. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun

Date 03/10/2022

(Narumon Worachun)

(Acting) Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society