



# NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

## CERTIFICATE OF ANALYSIS

**Product:** Low-Ionic Strength Salt Solution (LISS)

**Lot No.** 66011

**Mfg. Date** 27 JUL 23

**Volume** 10 mL

**Exp. Date** 27 JAN 25

Test Items	Specifications	Results
1. Appearance	Clear and colorless liquid	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.00	Complied
4. Conductivity	3.00 - 5.20 mS/cm	Complied
5. Osmolality	0.299 - 0.325 Osmol/kg	Complied
6. Serology test	LISS-IAT $\geq$ NSS-IAT	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun Date 02/11/2023

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society