

CERTIFICATE OF ANALYSIS

Product: Low-Ionic Strength Salt Solution (LISS)

Lot No. 67011

Mfg. Date 22 FEB 24

Volume 10 mL

Exp. Date 23 AUG 25

| Test Items | Specifications | Results |
|--------------------------|-----------------------------|----------|
| 1. Appearance | Clear and colorless liquid | Complied |
| 2. Label | Complete, Correct and Clear | Complied |
| 3. pH | 6.50 - 7.00 | Complied |
| 4. Conductivity | 3.00 - 5.20 mS/cm | Complied |
| 5. Osmolality | 0.299 - 0.325 Osmol/kg | Complied |
| 6. Serology test | LISS-IAT \geq NSS-IAT | Complied |
| 7. Microbial enumeration | NMT 1.0 CFU/mL | Complied |

Narumon Worachun Date 02/04/2024

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society