


NATIONAL BLOOD CENTRE  
THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

REAGENT: 3% Screening cells (O<sub>1</sub>, O<sub>2</sub>)  
LOT NO: 66117  
MFG DATE: 8 NOV 2023  
EXPIRE DATE: 15 DEC 2023  
SOURCE OF MATERIAL: Human Red Blood cells  
PRODUCT APPEARANCE: No hemolysis or turbid  
3% cells suspended in Preservative solution  
SPECIFICITY: Reagents cells must express the following  
Antigens: DCEce MN Mi<sup>a</sup> Ss P<sub>1</sub> Le<sup>a</sup> Le<sup>b</sup> Kk Fy<sup>a</sup> Fy<sup>b</sup> Jk<sup>a</sup> Jk<sup>b</sup> Di<sup>a</sup>  
STORAGE: Refrigerate of 2-8 °C

SIGNED..........DATE..... 9/11/2023 .....

Quality officer



พจนานุกรมศัพท์