



# NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

## CERTIFICATE OF ANALYSIS

**Product:** Anti-A

**Lot No.** 67012

**Source of Material:** Monoclonal antibody

**Mfg. Date** 20 MAR 24

**Volume** 10 mL

**Exp. Date** 20 SEP 25

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed blue No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	A <sub>1</sub> cell $\geq$ 1:256 A <sub>2</sub> B cell $\geq$ 1:64	Complied
5. Specificity test	A <sub>1</sub> cell = 4+ A <sub>2</sub> B cell >2+ B cell = Negative O cell = Negative	Complied
6. Avidity test	A <sub>1</sub> cell < 6 sec A <sub>2</sub> B cell < 30 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun Date 23/05/2024

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society