



NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

Product: Anti-A

Lot No. 65042

Source of Material: Monoclonal antibody

Mfg. Date 24 DEC 22

Volume 10 mL

Exp. Date 24 JUN 24

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed blue No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	A ₁ cell \geq 1:256 A ₂ B cell \geq 1:64	Complied
5. Specificity test	A ₁ cell = 4+ A ₂ B cell >2+ B cell = Negative O cell = Negative	Complied
6. Avidity test	A ₁ cell < 6 sec A ₂ B cell < 30 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun

Date 24/01/2023

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society