

CERTIFICATE OF ANALYSIS

Product: Anti-A	Lot No. 66023
Source of Material: Monoclonal antibody	Mfg. Date 07 MAY 23
Volume 10 mL, 50 mL	Exp. Date 07 NOV 24

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed blue No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	A ₁ cell ≥ 1:256 A ₂ B cell ≥ 1:64	Complied
5. Specificity test	A ₁ cell = 4+ A ₂ B cell >2+ B cell = Negative O cell = Negative	Complied
6. Avidity test	A ₁ cell < 6 sec A ₂ B cell < 30 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun Date 21/07/2023

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society