

CERTIFICATE OF ANALYSIS

Product: Anti-A,B

Lot No. 67011

Source of Material: Monoclonal antibody

Mfg. Date 05 JAN 24

Volume 10 mL, 50 mL

Exp. Date 05 JUL 25

Test Items	Specifications	Results
1. Appearance	Clear liquid dyed red No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	A ₁ cell ≥ 1:256 A ₂ cell ≥ 1:128 B cell ≥ 1:256	Complied
5. Specificity test	A ₁ cell = 4+ A ₂ cell = 4+ B cell = 4+ O cell = Negative	Complied
6. Avidity test	A ₁ cell < 6 sec A ₂ cell < 20 sec B cell < 6 sec	Complied
7. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun Date 06/02/2024

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society