

CERTIFICATE OF ANALYSIS

| | |
|--|----------------------------|
| Product: Anti-B | Lot No. 65031 |
| Source of Material: Monoclonal antibody | Mfg. Date 22 SEP 22 |
| Volume 10 mL, 50 mL | Exp. Date 22 MAR 24 |

| Test Items | Specifications | Results |
|--------------------------|--|----------|
| 1. Appearance | Clear liquid dyed yellow No turbidity, precipitate, particle or gel formation | Complied |
| 2. Label | Complete, Correct and Clear | Complied |
| 3. pH | 6.50 - 7.50 | Complied |
| 4. Titer | B cell \geq 1:256 A ₁ B cell \geq 1:128 | Complied |
| 5. Specificity test | B cell = 4+ A ₁ B cell = 4+ A ₁ cell = Negative O cell = Negative | Complied |
| 6. Avidity test | B cell < 6 sec A ₁ B cell < 6 sec | Complied |
| 7. Microbial enumeration | NMT 1.0 CFU/mL | Complied |

Narumon Worachun Date 21/10/2022

(Narumon Worachun)

(Acting) Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society