



NATIONAL BLOOD CENTRE, THAI RED CROSS SOCIETY

CERTIFICATE OF ANALYSIS

Product: Anti-c

Lot No. 65020

Source of Material: Monoclonal antibody

Mfg. Date 15 NOV 22

Volume 5 mL

Exp. Date 15 MAY 24

Test Items	Specifications	Results
1. Appearance	Clear liquid No turbidity, precipitate, particle or gel formation	Complied
2. Label	Complete, Correct and Clear	Complied
3. pH	6.50 - 7.50	Complied
4. Titer	R ₁ r cell ≥ 1:32	Complied
5. Specificity test	R ₁ r cell ≥ 2+ R ₁ R ₂ cell ≥ 2+ r'r cell ≥ 2+ R ₁ R ₁ cell = Negative	Complied
6. Microbial enumeration	NMT 1.0 CFU/mL	Complied

Narumon Worachun Date 29/11/2022

(Narumon Worachun)

Quality Assurance and Quality Control Manager

National Blood Centre, Thai Red Cross Society